

A Comparison of Labsocs and Source-Based Full Energy Efficiency Generation in Measurements

***R.L. Njinga, and V.M. Tshivhase**

Center for Applied Radiation Science and Technology, North-West University,
Mafikeng, South Africa

njingaraymond@yahoo.co.uk

Abstract: *Not all gamma rays emitted by the source that pass through the detector will produce a count in the system. The probability that an emitted gamma ray will interact with the detector and produce a count is the efficiency of the detector and is measured by comparing a spectrum from a source of known activity, count rates in each peak, count rates expected from the known intensities of each gamma ray. The energy of the gamma rays being detected is an important factor in the efficiency of the detector. An efficiency curve can be obtained by plotting the efficiency at various energies. This curve can then be used to determine the efficiency of the detector at energies different from those used to obtain the curve. A comparison of two gamma-ray efficiency determination methods; Canberra's LabSOCS and the source-based efficiency calibrations using marinelli beaker geometry were measured and the results revealed +96% confidence levels for each gamma peak lines. This means that the Canberra's LabSOCS will be used for efficiency curve generation and no more dependence on standard sources in activity concentration measurements for variety of sample matrices.*

Keywords: *LabSOCS, efficiency, energy, standard sources, marinelli beaker, efficiency curves*

1. INTRODUCTION

Although high purity germanium detectors (HPGe) detectors come in many different designs and sizes, the most common type of detector is the coaxial detector which in centre for applied science and technology (CARST) is useful for measurement of gamma-rays with energies over the range from about 31.00 keV to 2.50 MeV. Laboratories which are directly involved in the measurement and certification of the radioactive content of various samples collected from the environment and the food chain, or industrial products should development analysis procedures to qualitatively or quantitatively evaluate the radionuclides concentration involved [1]. This involves set-up for calibrations (energy and efficiency) development of the HPGe detectors for each measurements to be employed depending on the varieties of sample sizes (geometries) with different densities [2]. Moreover to take full advantage of the large number of radionuclides whose concentration are determined from the main natural gamma-ray emitters, the efficiency should be known at least from 31.00 keV to 2.50 MeV, associated with the decay products of the ^{238}U and ^{232}Th decay series [3].

In order to perform accurate and reliable activity measurements, periodical calibrations of the gamma-ray spectrometry installations of the laboratories are required [4]. In accredited laboratories, the gamma-ray spectrometry method is used to perform both qualitative and quantitative radioactivity analysis, for solid, liquid and gaseous samples [5, 6]. A typical gamma-ray spectrometry system is composed of; a detector usually semiconductor, such as HPGe with a shielding - mainly lead, to reduce the background; high voltage power supply; electronics for signal processing (preamplifier, amplifier, multichannel analyzer); computer and dedicated software [7].

The spectrometric system actually records, stores and processes the gamma-ray spectrum of any analysed sample, using validated computer software packages [8, 9]. This means that a proper energy and efficiency calibration is needed [10] to identify the energy of the gamma ray emissions from the spectrum or the gamma-ray emitter radionuclides contained by the sample (qualitative analysis). The quantitative analysis or the activity and its standard uncertainty determination for each of the radionuclide present in the sample, requires a full-energy peak efficiency (FEPE) calibration [11]. For energy and source based efficiency calibrations, various radioactive standard sources with certified activity are required.

The high resolution HPGe detector in CARST is suitable for samples containing many radionuclides from the natural radioactive series when the gamma-ray spectrum presents a large number of peaks to be resolved. These detectors are usually kept at very low temperatures (in liquid nitrogen) for a correct functioning. The detectors perform faster in analysis due to the high efficiency advantage.

This work is aimed at obtaining the energy calibration and parameters of the second degree polynomial equation and comparison of the Laboratory Sourceless Calibration Software (LabSOCS) and source-based efficiency curves generated. It also investigates the differences in efficiency curves generated due to variation of source heights by ~ 1.76% and ~10.07% from the actual source height using the simulated LabSOCS.

2. EXPERIMENT

The energy calibration implies the experimental determinations of a function of either first or second degree polynomial, describing the energy dependence of the channel number in the spectrum as:

$$E_{\gamma} = a + b \times ch + c \times ch^2 \quad 1.0$$

Where E_{γ} is the gamma-ray energy, ch is the channel with the maximum number of counts, $a, b,$ and c are constants to be determined for calibration.

The energy calibration is performed manually (cursor, marker) or automatically by the software, by measuring one or several radioactive standard sources emitting gamma-rays of minimum of three to seventeen different energies covering a wide spectral range between 31.00 keV and 2.50 MeV.

In the approach of efficiency generation using LabSOCS, the materials have been developed using the geometry composer feature of Canberra's Genie™ 2000 Version 2.0 and Gamma Analysis Version 2.0A software packages. The container matrix material for the Marinelli beaker was simulated for polypropylene with chemical composition of 14.37% H, 85.63% C and 100% C₃H₆ from the materials library file with a default density of 0.91 g/cc. As shown in Table 1.0, the customized beaker files were created to accurately define the inner and outer wall contours, the material, and density value of the container. These files were created using a standard text editor and stored as one of the "beaker" templates in the geometry composer window to represent the Marinelli beakers available in CARST.

Table1. Dimension entered in simulating the Marinelli beaker and soil-like source

Description	d.1 (cm)	d.2(cm)	d.3(cm)	d.4 (cm)	Material	Density (g/cc)
Container	0.25	1.35	9.00	7.50	polypropylene	0.91
Source heights	d ₁ , d ₂ , d ₃	-	-	-	Soil	1.40
Source-detector	0.15	-	-	-	-	-

$10.04\text{cm}=d_1$ (LabMBkr1.ECC), $10.22\text{cm}=d_2$ (LabMBkr2.ECC), $11.25\text{cm}=d_3$ (LabMBkr3.ECC)

Table 1 shows a detailed description of the parameter values used to define the dimensions and material composition of the container, sample matrix and source-to-detector distance. For the sample model described, the LabSOCS Version 4.0 software was used to generate a set of mathematically calculated efficiency values for a specified set of energy values entered for 31.00 keV (¹³³Ba) to 1408.01 keV (¹⁵²Eu) with appropriate corresponding percentage uncertainty values ranging from 10% at low energies to 4% at high energies. A customized energy list created was stored as a text file. The detector of diameter (6.21 cm) with characterization file was created and used for the LabSOCS modelling and efficiency calculations.

From the menu bar, the Edit/Efficiency/parameter; Edit/Efficiency/validate geometry and Edit/Efficiency/generate data point options were followed sequentially. The efficiency data were generated for each of the source heights 10.04 cm (d₁), 10.22 cm (d₂) and 11.25 cm (d₃) respectively as shown in Fig. 1. The Gamma Acquisition and Analysis (GAA) window was launched, and a pre-existing CAM file data source opened in the GAA window.

The Calibrate|Efficiency|By Entry option was selected from the GAA window menu bar and each efficiency data with the respective uncertainty values were entered in the displayed dialog box opened and the 'show' action button was used to display the Dual, Empirical and Linear efficiency curves. The orders of the polynomial for the efficiency curve types were modified as necessary to achieve the best curve fits. The 'report' action button was used to generate a one-page report of the LabSOCS efficiency results. The "Calibrate/Store action button was used to save the results as a standard Genie

2000 efficiency calibration file in the Genie2k\Calfiles folder and the finish action button was used to close the dialog box to return to the GAA window. The CALfiles created were then opened as a CAM file data-source in the GAA window and the Calibrate|Efficiency and the ‘show’ option was selected from the menu bar, followed by the ‘print action’ button to generate printed plots of the appropriate efficiency curve. From the reports generated, each of the data values were manually copied into the Microsoft excel spread sheet for further analysis.




Figure1. Demonstration of source heights (d1, d2, and d3)

For the source-based, the efficiency data were generated using the radioactive standard sources (^{133}Ba and ^{152}Eu) with gamma-ray emissions covering the wide energy range. For a given peak corresponding to the energy in the gamma-ray spectrum measured with the HPGe detector, the efficiency ε and its expanded standard uncertainty were computed as follows [12]:

$$\varepsilon = \frac{(N_{PA} - N_{BPA}) \times F_D}{T \times G_I \times A_{STD} \times F_C \times F_T} \quad 2.0$$

where N_{PA} and N_{BPA} are the net areas of the considered peak from the gamma-ray spectrum of the in house soil standard and the net background peak area of the respective spectrum generated, T is the measurement time (the same for standard and background, expressed in seconds), F_C , F_T and F_D are the multiplicative coefficients for coincidence summing corrections, efficiency transfer corrections and for the decay during the reference time and the time the sample measurement started (important only for short half-life radionuclides), G_I is the emission probability of the considered gamma-rays and A_{STD} is the activity of the in house soil standard, expressed in Becquerel. The uncertainty in using equation 2.0 was evaluated as follows:

$$U_\varepsilon = \sqrt{S_{N_{PA}-N_{PBA}}^2 + U_{G_I}^2 + U_{A_{STD}}^2 + U_{F_D}^2 + U_{F_C}^2 + U_{F_T}^2 + U_T^2} \quad 3.0$$

Where U_ε is the efficiency combined standard uncertainty, $S_{N_{PA}-N_{PBA}}^2$ is the result of the net peak area (background subtracted) counts measurement uncertainty, while $U_{G_I}^2$, $U_{A_{STD}}^2$ and $U_{F_C}^2$, are the results of the measurement uncertainties for the gamma emission probability, activity and correction factors for coincidence summing, $U_{F_T}^2$ is measurement uncertainties for efficiency transfer, $U_{F_D}^2$ is measurement uncertainties for decay during the reference time and the time of the measurement start and U_T^2 is measurement uncertainties of the time. However, the standard uncertainty of the measurement time and the corresponding decay are considered negligible (=1) in this work.

The mixed-gamma (^{133}Ba and ^{152}Eu) standard (soil matrix form) in the 1 litre marinelli beaker was counted until satisfactory counts (approximately 20,000) in each certificate peak criteria were recorded [13]. The efficiency calibration calculations were performed using the GAA Calibrate|Efficiency|By Certificate File option from the GAA menu bar and the report printed. From the reports printed, all the data points corresponding to the gamma peak lines from the certificate file were manually transferred to the Microsoft excel spread sheet for comparisons.

3. RESULT AND DISCUSSIONS

Form the energy versus channel calibration, we obtained the relationship with the values of parameters; a, b, and c from the equation (Energy = 0.7246 + 3.0259 ch + 2.0 x 10⁻⁰⁶ch²) as showed in Figure 2.0. The linear relationship following the regressive line showed a value of R² to be 1.00 with 100% confidence.




Figure2. Channels versus energy calibration

The efficiency curves generated were limited to the geometries of identical matrices and different source heights as shown in Figure 3.0. The comparison was related to the source height of 10.22 cm (d_2). The efficiency curves obtained from the LabSOCS and the source base (Figure 3.0) shows that, almost all the data points agreed with +96.3% confidence intervals.




Figure3. Efficiency curve at d_1 , d_2 , d_3 and the mixed source base standard

This confirms that the LabSOCS efficiency calibration technique produces efficiency values which agree with source-based efficiency calibrations for the marinelli beaker containers. Therefore, employing LabSOCS for efficiency curves generation will be adopted in CARST and this will reduce costs associated with purchase, maintenance and disposal of radioactive standard sources. In addition, the LabSOCS technique, using the ‘geometry composer’, will enable researchers to produce assay-grade measurements of unique sample/matrix/container samples such as water, soil, gravel, sediments, sand and certain biological samples faster.

Also, the efficiency curves of the different source heights obtained from LabSOCS are shown in Fig. 3. From the curves, it is evidence that slight misjudgements of source height from the actual, will either reduced by ± 1.4 to ± 2.79 % or increased by ± 8.69 to ± 16.32 %. Hence, homogeneity in sample matrix (sample texture), sample heights and container types, counting time and other physical parameters must be adhered to in order to obtained credible results.

4. CONCLUSION

This work describes a comparison of two gamma-ray efficiency determination methods; - Canberra’s LabSOCS (Laboratory SOURCEless Calibration Software) and the source-based efficiency calibrations using marinelli beaker sample container geometry. LabSOCS geometry modelling was developed using the geometry composer feature of Canberra’s Genie™ 2000 version 2.0 and Gamma Analysis version 2.0A software packages. The efficiency data generated for each source height 10.04 cm, 10.22 cm, and 11.25cm shows variation with respect to the actual source height 10.22 cm. Based on the

actual height of 10.22 cm, the LabSOCS efficiencies compared well with the source-based efficiencies for the marinelli beaker with confidence level of +96% for each of the energy peak lines.

REFERENCES

- [1] Park, T.S., Jeon, W.J., 1995. Measurement of radioactivity samples in Marinelli beakers by gamma- ray spectrometry. *J. Radioanal. Nucl. Chem.* 193, 133–144.
- [2] F. Bronson & R. Venkataraman, “Validation of the Accuracy of the LabSOCS Mathematical Efficiency Calibration for Typical Laboratory Samples”. 46th Annual Conference on Bioassay, Analytical and Environmental Radiochemistry, November 11-17, 2000, Seattle, WA.
- [3] M. J. Daza, B. Quintana M. Garca Talavera, F. Fernandez “Efficiency calibration of a HPGe detector in the [46.54-2000] keV energy range for the measurement of environmental samples” *Nuclear Instruments and Methods in Physics Research A* 470 (2001) 520-532
- [4] “ANSI Standard for Calibration and Use of Germanium Spectrometers for the Measurement of Gamma-Ray Emission Rates of Radionuclides”, ANSI N42.14-1999, Table 2, Page 22.
- [5] N. Lavia, Z. B. Alfassi, “Development and application of Marinelli beaker standards for monitoring radioactivity in Dairy-Products by gamma-ray spectrometry” *Applied Radiation and Isotopes* 61 (2004) 1437–1441
- [6] Taskin H, Karavus M, Topuzoglu A, Hindiroglu S, Karahan G (2009). Radionuclide concentrations in soil and Lifetime cancer risk due to the gamma radioactivity in Kirklareli, Turkey. *J. Environmental Radioactivity.* 100: 49-53.
- [7] P. W. Gray, A. Ahmed “Linear classes of Ge(Li) detector efficiency functions” *Nuclear Instruments and Methods in Physics Research A* 237, 577-589, (1985).
- [8] J. Saegusa, Katsuya Kawasaki, Akira Mihara, Mitsuo Ito, Makoto Yoshidaa "Determination of detection efficiency curves of HPGe detectors on radioactivity measurement of volume samples" *Applied Radiation and Isotopes* 61 (2004) 1383–1390
- [9] M. Jurado Vargas, A. Fernandez Timon, N. Cornejo Diaz, D. Pérez Sánchez “Monte Carlo simulation of the self-absorption corrections for natural samples in gamma-ray spectrometry” *Applied Radiation and Isotopes* 57 (2002) 893–898
- [10] D. Groff, “LabSOCS Geometry Modeling and Efficiency Calibration File Generation for TVA Sample Fixtures”, December 6, 2001
- [11] M. I. Abbas, Younis S. Selim “Calculation of relative full-energy peak efficiencies of well-type detectors” *Nuclear Instruments and Methods in Physics Research A* 480 (2002) 651–657
- [12] G. Hasse, D. Tait and A. Wiechen “determination of full energy peak efficiency for cylindrical volume sources by the use of a point source standard in gamma-spectrometry” *Nuclear Instruments and Methods in Physics Research A* 361 (1995) 240-244
- [13] Z. B. Alfassi, N. Lavi “The dependence of the counting efficiency of Marinelli beakers for environmental samples on the density of the samples” *Applied Radiation and Isotopes* 63 (2005) 87–92